



Indian Institute of Technology Bhilai (IIT Bhilai)
GEC Campus, Old Dhamtari Road, Sejbahar, Raipur, Chhattisgarh, India – 492015
www.iitbhilai.ac.in

NOTICE INVITING TENDER

Indian Institute of Technology Bhilai (IIT Bhilai), an autonomous body under the Ministry of Education, Govt of India invites e-bid for **“Supply and Installation of High Temperature Tribometer at IIT Bhilai”**. Tender Documents may be downloaded from Central Public Procurement Portal ([URL:http://eprocure.gov.in/eprocure/app](http://eprocure.gov.in/eprocure/app)). Bidders are advised to submit duly filled bids as per the following schedule:

Invitation for Bids (IFB) No & date	IITBh/Goods/Mech/2021-22/220 Dated: 13/01/2022
Place of Supply etc.	Indian Institute of Technology Bhilai GEC Campus, Old Dhamtari Road, Sejbahar, Raipur-492015, Chhattisgarh
Pre-bid Meeting date	January 20, 2022, 02: 30 PM (through WebEx) Any query (including technical) related to pre-bid meeting, mail should be reached on or before 20/01/2022 01:00 PM. Interested bidders can request the WebEx meeting link at sp@iitbhilai.ac.in
Bid Submission Start date	January 27, 2022, 1100 Hrs
Last Date of Submission of Bids	February 8 , 2022, 1500 Hrs
Date of Opening of Technical Bids	February 9, 2022, 1515 Hrs
Tender Fee & EMD	Tender Fee - Rs. 1500/- and EMD – Rs.2,60,000/- should be paid using SBI i-collect Link https://www.onlinesbi.com/sbicollect/icollecthome.htm?corpID=646425 Step1. Click Check Box to proceed for payment. Step2. Select S and P Tender Fee and EMD. Step3. Fill all Details and Submit (This online payment receipt may be provided in the on line Technical e-bid) EMD - MSE's are exempted from the payment of EMD, MSME's should submit the Valid MSME certificate in manufacturing of Tribometer . MSMEs should submit the Bid Security Declaration (in Annexure-H).
Place of Opening of Technical Bids	Online - E-procurement Portal
Contact information (Tender Inviting Authority)	Administrative Advisor Indian Institute of Technology Bhilai, GEC Campus, Old Dhamtari Road, Sejbahar, Raipur 492015, Chhattisgarh Phone: +91-771-2973622, Email: sp@iitbhilai.ac.in For technical enquiries: Email: purnendu@iitbhilai.ac.in

Supply and Installation of High Temperature Tribometer at IIT Bhilai

CALL INSTRUCTIONS FOR ONLINE BID SUBMISSION

The bidders are required to submit their bids electronically on the CPP Portal using valid Digital Signature Certificates. Certain instructions are given below to guide the bidders for registration on the CPP Portal, prepare their bids in accordance with the requirements and submitting their bids online on the CPP Portal.

Registration Process

- 1) If a bidder is not registered at the eProcurement portal, such bidder should enrol/register on the e-Procurement module of the portal <http://eprocure.gov.in/eprocure/app> by clicking on the link "Click here to Enrol". Enrolment on the CPP Portal is free of charge.
- 2) The bidders to choose a unique username and assign a password for their accounts. Bidders are advised to register their valid email address and mobile numbers as part of the registration process. These would be used for any communication from the CPP Portal.
- 3) Bidders to register upon enrolment their valid Digital Signature Certificate (Class II or Class III Certificates with signing key usage) issued by any Certifying Authority recognized by CCA India with their profile.
- 4) Only one valid DSC should be registered by a bidder. Please note that the bidders are responsible to ensure that they do not lend their DSCs to others which may lead to misuse. *Foreign bidders are advised to refer "DSC details for Foreign Bidders" for Digital Signature requirements on the portal.*
- 5) Bidder then logs in to the site through the secured log-in by entering their user ID / password and the password of the DSC / eToken.

Tender Documents Search

- 1) Various built in options are available in the CPP Portal to facilitate bidders to search active tenders by several parameters. These parameters include Tender ID, organization, location, date, value, etc.
- 2) There is also an option of advanced search for tenders, wherein the bidders may combine a number of search parameters such as organization name, form of contract, location, date, other keywords etc. to search for a tender published on the CPP Portal.
- 3) Once the bidders have selected the tenders they are interested in, they may download the required documents / tender schedules. These tenders can be moved to the respective 'My Tenders' folder. This would enable the CPP Portal to intimate the bidders through SMS / e-mail in case there is any corrigendum issued to the tender document.
- 4) The bidder should make a note of the unique Tender ID assigned to each tender, in case they want to obtain any clarification / help from the Helpdesk.

Bid Preparation

- 1) Bidder should also take into account of the corrigendum published related to the tender before submitting their bids.
- 2) Please go through the tender advertisement and the tender document carefully to understand the documents required to be submitted as part of the bid.
- 3) Please note the number of covers in which the bid documents have to be submitted, the number of documents - including the names and content of each of the document that need to be submitted. Any deviations from these may lead to rejection of the bid.
- 4) Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document / schedule and generally, they can be in PDF / XLS / RAR / DWF formats. Bid documents may be scanned with **100 dpi with black and white option**.
- 5) To avoid the time and effort required in uploading the same set of standard documents which are required to be submitted as a part of every bid, a provision of uploading such standard documents (e.g. PAN card copy, annual reports, auditor certificates etc.) has been provided to the bidders. Bidders can use "My Space" area available to them to upload such documents. These documents may be directly submitted from the "My Space" area while submitting a bid, and need not be uploaded again and again. This will lead to a reduction in the time required for the bid submission process.

Bid Submission

Supply and Installation of High Temperature Tribometer at IIT Bhilai

- 1) Bidder to log into the site well in advance for bid submission so that he/she uploads the bid in time i.e. on or before the bid submission time. Bidder will be responsible for any delay due to other issues.
- 2) The bidder to digitally sign and upload the required bid documents one by one as indicated in the tender document.
- 3) Bidder to select the payment option as "Off-line" to pay the tender fee/ EMD wherever applicable and enter details of the instrument in SBI i-collect.
- 4) A standard BoQ format has been also uploaded along with the tender document to be filled by all the bidders. Bidders to note that they should necessarily submit their financial bids in the prescribed format and no other format is acceptable.
- 5) The server time (which is displayed on the bidders' dashboard) will be considered as the standard time for referencing the deadlines for submission of the bids by the bidders, opening of bids etc. The bidders should follow this time during bid submission.
- 6) All the documents being submitted by the bidders would be encrypted using PKI encryption techniques to ensure the secrecy of the data, which cannot be viewed by unauthorized persons until the time of bid opening.
- 7) The uploaded tender documents become readable only after the tender opening by the authorized bid openers.
- 8) Upon the successful and timely submission of bids, the portal will give a successful bid submission message & a bid summary will be displayed with the bid no. and the date & time of submission of the bid with all other relevant details.
- 9) Kindly add scanned PDF of all relevant documents in a single PDF file of compliance sheet.

Assistance to Bidders

- 1) Any queries relating to the tender document and the terms and conditions contained therein should be addressed to the Tender Inviting Authority for a tender or the relevant contact person indicated in the tender.
- 2) Any queries relating to the process of online bid submission or queries relating to CPP Portal in general may be directed to the 24x7 CPP Portal Helpdesk. The contact number for the helpdesk is **0120-4200462, 0120-4001002, 0120-4001005**.

General Instructions to the Bidders

- The tenders will be received online through the portal <https://eprocure.gov.in/eprocure/app>. In the Technical Bids, the bidders are required to upload all the documents in .pdf format.
- Possession of Valid Class II/III Digital Signature Certificate (DSC) in the form of smart card/ e-Token in the company's name is a prerequisite for registration and participating in the bid submission activities through <https://eprocure.gov.in/eprocure/app>. Digital Signature Certificates can be obtained from the authorized certifying agencies, details of which are available in the web site <https://eprocure.gov.in/eprocure/app> under the link 'Information about DSC'.

Tenderers are advised to follow the instructions provided in the 'Instructions to the Tenderers for the e-Submission of the bids online through the Central Public Procurement Portal for e-Procurement at <https://eprocure.gov.in/eprocure/app>.

LIST OF DOCUMENTS TO BE UPLOADED IN E-PROCUREMENT PORTAL

- 1) Bidder should take into account notifications, corrigendum published, if any on the tender document before submitting their bids.
- 2) Please go through the tender advertisement and the tender document carefully to understand the documents required to be submitted as part of the bid. Any deviations from these may lead to rejection of the bid.
- 3) The two-bid system will be followed for this tender. In this system, online offer should be submitted under **TWO-BID System** in two separate e-packets i.e. **"Technical eBid" and "Commercial eBid"**.

Online Envelope No. 1: "Technical Bid" shall contain (Pdf format only)

Supply and Installation of High Temperature Tribometer at IIT Bhilai

1. Scanned copy of tender fee and EMD Payment.
2. EMD is exempted for Micro and Small Enterprises (MSE) which are authorized in manufacturing of **Tribometer**. Micro and Small Enterprises (MSE) as defined in MSE Procurement Policy issued by the Department of Micro, Small and Medium Enterprises (MSME) / NSIC and Start-ups as recognised by Department of Industrial Policy & Promotion (DIPP), Ministry of Commerce and Industry.
MSMEs shall submit Bid Security Declaration (Refer Format attached as Annexure - H.) The Exemption Certificate / Declaration should be enclosed in Technical Bid Only.
Those firms, who are failing to submit the Valid MSME or Bid Security Declaration (As per annexure-H), their bid will be summarily rejected.
3. A scanned copy of Authorization letter with proper seal and signature of the authorised person (with name, designation, email id & contact no.), as per the format attached at Annexure – C.
4. A scanned copy of Certificate of Incorporation, Partnership Deed/ Memorandum and Articles of Association/ any other equivalent document showing date and place of incorporation, as applicable.
5. A scanned copy of the commercial bid without prices (prices blocked) and copy of commercial terms & conditions (in details) as included in the commercial bid. IIT Bhilai reserves the right to reject the bid in case of discrepancy observed in the unpriced commercial bid and the actual commercial bid.
6. A scanned copy of audited balance sheet or the certificate from a Chartered Accountant for the financial year 2020-21, 2019-20 and 2018-19 indicating the annual sales turnover of the bidder. Bidder firm should not have suffered any financial loss for more than one year during the last three years, ending on 'The Relevant Date'.
7. The detailed technical specification, make, model, part number & compliance of each item offered supported by the printed catalogue/ leaflet published by the Principal Manufacturer.
8. Undertaking to the effect that a Performance Guarantee of 3% of the order value will be submitted in case IIT Bhilai decides to place the Purchase Order.
9. Undertaking to the effect that the terms and conditions, clauses etc. stipulated in this tender are acceptable.
10. Other documents necessary in support of eligibility criteria (as per Section-II - Instructions to Bidders-ITB), product catalogues, brochures etc.

IIT Bhilai reserves the right to reject the bid if any of the above listed document/s is not submitted.

Online Envelope 2: "Commercial Bid" shall contain

- 1) The Commercial Bid should be filled properly in the .xls uploaded in eProcurement portal
- 2) In case the bidder requires any clarifications/ information, they may write to sp@iitbhilai.ac.in.
- 3) Commercial bids of the technically qualified bidders only will be opened in the eProcurement portal.
- 4) Commercial bids must be offered in the format attached.
- 5) The applicable exact rates of Taxes will be added as per the format attached.
- 6) In case of any mistake or error in calculations or any discrepancy in price quoted in words and figures, the LOWEST amount will be considered for comparison. If this lowest price is not acceptable to the bidder, their bid will be rejected. In this case, IIT Bhilai reserves the right to forfeit their EMD.
- 7) IIT Bhilai reserves the right to place part orders.

Supply and Installation of High Temperature Tribometer at IIT Bhilai

SECTION-I: INSTRUCTIONS TO BIDDERS (ITB)

1. Locations for the Supply and Installation & Warranty Services

The entire products as described in Schedule of Requirements at **Section - V** must be supplied, installed & supported at Indian Institute of Technology, Bhilai.

2. All the items covered in the Schedule of Requirements at Section – V must be supplied within 5 months from the date of placement of work/ purchase order. Supplier should complete the installation within 30 days from the date of delivery of the equipment.

3. Amendment to Bidding Documents

- 3.1 At any time prior to the deadline for submission of bids, IIT Bhilai may, for any reason, whether on its own initiative or in response to the clarification request by a prospective bidder, modify the bid document.
- 3.2 The amendments to the tender documents, if any, will be notified by release of Corrigendum Notice on the institute website against this tender. The amendments/ modifications will be binding on the bidders.
- 3.3 IIT Bhilai at its discretion may extend the deadline for the submission of bids if it thinks necessary to do so or if the bid document undergoes changes during the bidding period, in order to give prospective bidders time to take into consideration the amendments while preparing their bids.

4. Preparation of Bids

Bidder should avoid, as far as possible, corrections, overwriting, erasures or postscripts in the bid documents. In case however, any corrections, overwriting, erasures or postscripts have to be made in the bids, they should be supported by dated signatures of the same authorized person signing the bid documents. However, bidder shall not be entitled to amend/ add/ delete/ correct the clauses mentioned in the entire tender document.

5. Forfeiture of EMD

- 5.1 In case of the bidder whose offer is accepted, the EMD will be returned on submission of Security Deposit. However, if the return of EMD is delayed for any reason, no interest/ penalty shall be payable to the bidder.
- 5.2 The successful bidder, on award of contract/ order, must send the contract/ order acceptance in writing, within 15 days of award of contract/ order, failing which the EMD will be forfeited and the order will be cancelled.
- 5.3 If the bidder withdraws the bid during the period of bid validity specified in the tender.

6. Period of validity of bids

- 6.1 Bids shall be valid for minimum 180 days from the date of submission. A bid valid for a shorter period shall stand rejected.
- 6.2 IIT Bhilai may ask for the bidder's consent to extend the period of validity. Such request and the response shall be made in writing only. The bidder is free not to accept such requests without forfeiting the EMD/ BG. A bidder agreeing to the request for extension will not be permitted to modify his bid.

7. Deadline for Submission of Bids – online only

- 7.1 The Bid documents shall be neatly arranged. They should not contain any terms and conditions, printed or otherwise, which are not applicable to the Bid. *The conditional bid will be summarily rejected.*
- 7.2 Bids must be received by IIT Bhilai before the due date and time at the address specified in the tender document.
- 7.3 IIT Bhilai may extend this deadline for submission of bids by amending the bid documents and the same shall be suitably notified on its website.
- 7.4 IIT Bhilai shall not be responsible and liable for **the delay in receiving the bid** for whatsoever reason.

8. Bid Opening & Evaluation of Bids

- 8.1 The technical bids will be evaluated in two steps.

Supply and Installation of High Temperature Tribometer at IIT Bhilai

- a. The bids of eligibility bidders only will be evaluable based on eligibility criteria mentioned in Section III of the tender documents.
 - b. The technical bids of only the shortlisted eligible bidders shall be evaluated based on technical specifications stipulated at Section – V
- 8.2 The bidders whose technical bid is found to meet both the requirements as specified above will qualify for opening of the commercial bid.
 - 8.3 The duly constituted Tender Evaluation Committee (TEC) shall evaluate the bids. The TEC shall be empowered to take appropriate decisions on minor deviations, if any.
 - 8.4 The bidder's name, bid prices, discounts and such other details considered as appropriate by IIT Bhilai, will be announced at the time of opening of the commercial bids.

9. Comparison of Bids

- 9.1 Only the short-listed bids from the technical evaluation shall be considered for commercial comparison.
- 9.2 The bids shall be evaluated on the basis of total price quoted for Sl. No. 1, Supply and installation of High Temperature Tribometer at IIT Bhilai in the Commercial bid.
- 9.3 The bids shall be evaluated on the basis of the total prices including all taxes and duties as per the details mentioned in section I (10.2) and section VI (8).
- 9.4 In the comparison of bids, preference will be given to Class – I Local Suppliers as per the provisions of the Order no. P-45021/2/2017-PP (BE-II) on Public Procurement (Preference to Make in India), 2017 dated 04.06.2020 and subsequent orders issued by Department for Promotion of Industry and Internal Trade. Provisions of the order can be referred from the link given below:
<https://dipp.gov.in/sites/default/files/PPP%20MII%20Order%20dated%204th%20June%202020.pdf>

10. Purchase Order

- 10.1 IIT Bhilai shall issue purchase order to the eligible bidder whose technical bid has been accepted and determined as the lowest evaluated commercial bid based on the Grand Total calculated of all items as mentioned in section VI(8) + taxes of the Commercial Bids. However, IIT Bhilai reserves the right and has sole discretion to reject the lowest evaluated bid.
- 10.2 If more than one bidder happens to quote the same lowest price, IIT Bhilai reserves the right to decide the criteria and further process for awarding the contract, decision of IIT Bhilai shall be final for awarding the contract.

11. End User Certificate (if applicable)

IIT Bhilai can provide end user certificate in advance in order to save on time/ delivery period. However, this does not mean that the order shall be placed on the same bidder. The bidder may submit the format of end user certificate at the earliest, preferably along with the technical bid.

12. Purchaser's Right to amend / cancel

- 12.1 IIT Bhilai reserves the right to amend the eligibility criteria, commercial terms & conditions, Scope of Supply, technical specifications etc.
- 12.2 IIT Bhilai reserves the right to cancel the entire tender without assigning any reasons thereof.
- 12.3 IIT Bhilai reserves the right to place part orders.

13. Corrupt or Fraudulent Practices

- 13.1 It is expected that the bidders who wish to bid for this project have the highest standards of ethics.
- 13.2 IIT Bhilai will reject bid if it determines that the bidder recommended for award has engaged in corrupt or fraudulent practices while competing for this contract.
- 13.3 IIT Bhilai may declare a vendor ineligible, either indefinitely or for a stated duration, to be awarded a contract if it at any time determines that the vendor has engaged in corrupt and

Supply and Installation of High Temperature Tribometer at IIT Bhilai
fraudulent practices during the award / execution of contract.

14. Interpretation of the clauses in the Tender Document / Contract Document

In case of any ambiguity/ dispute in the interpretation of any of the clauses in this Tender Document, the interpretation of the clauses by Director, IIT Bhilai shall be final and binding on all parties.

***** (END OF SECTION I) *****

Supply and Installation of High Temperature Tribometer at IIT Bhilai

SECTION-II: ADDITIONAL INSTRUCTIONS TO BIDDERS (AITBs)

- (a) Delivery: The items be supplied within 5 months from the date of receipt of the Purchase Order.
- (b) Installation: Installation should be completed within 30 days from the date of supply or as per the instructions of IIT Bhilai, whichever is later
- (c) Bidder should also submit the PO copies of similar tribometer installations in India.
- (d) Authorization certificate from all the original equipment manufacturers stating their support for the vendor during the equipment warranty period. Warranty of each component should be mentioned separately, if applicable.
- (d) The Competent Authority reserves the right to reject any or all bids, amend/withdraw any terms and conditions in the Tender document or to cancel the Tender at any stage without assigning any reason.**

***** (END OF SECTION II) *****

Supply and Installation of High Temperature Tribometer at IIT Bhilai

SECTION-III: ELIGIBILITY AND QUALIFICATION CRITERIA

1. The bidder must be either OEM or authorized dealers or resellers of the original manufacturer of the proposed equipment in India. Both cannot bid simultaneously for the same item/ product in the same tender.
2. The vendor must have supplied at least similar equipment to at least 5 educational institutes/universities and/or research organizations in India and/or abroad along with email addresses, where the similar systems have already been installed. The similar equipment must bear the nominal specifications (For example: Temperature 1000 Degree C or above and load range: Max. Load (N): 50 or Higher and Min. Load (N): 1 or Lower and all other specifications mentioned in the current tender).
Complete list along with the contact details of the customers must be provided. Feedback from the users will be sought by the IIT Bhilai, if felt necessary. They also need to submit the POs as proof.
3. A) In addition to the above eligibility criteria, bidders should fulfil the minimum local content eligibility criteria as per the provisions of the Order no. P-45021/2/2017-PP (BE-II) on Public Procurement (Preference to Make in India), 2017 dated 04.06.2020 issued by Department for Promotion of Industry and Internal Trade.

B) Local content calculation should be as per the provisions of the OM No. P-45021/102/2019-BE-II-Part (1) (E-50310) Dated: 04-03-2021 issued by Department of Promotion of Industry and Internal Trade, Ministry of Commerce and Industry.

C) Bidders' eligibility should also be as per the subsequent orders on local content or purchase preference issued by the Department for Promotion of Industry and Internal Trade in due course.
4. For the purpose of claiming purchase preference, the bidders should have the minimum local content of at least 50% and margin of preference shall be 20%. Bidders should submit a self-certification (as per annexure -G) that the item meets the minimum local content and shall give details of such local value addition.
5. If in the view of bidder, any exemption/ relaxation is applicable to them from any of the eligibility requirements, under any Rules/ process/ Guidelines/ Directives of Government of India, bidder may submit their claim for the applicable exemption/ relaxation, quoting the valid Rule/ process/ Guidelines/ Directives. In this case the bidder must submit necessary and sufficient documents along with the technical bid, in support of his claim. The applicability of exemptions and /or validity of documents submitted by the bidders will be evaluated by the bid evaluation committee during technical evaluation of bids.

Note: *The bidders should provide sufficient documentary evidence to support the eligibility Criteria. IIT Bhilai reserves the right to reject any bid not fulfilling the eligibility criteria.*

***** (END OF SECTION III) *****

Supply and Installation of High Temperature Tribometer at IIT Bhilai

SECTION-IV: GENERAL CONDITIONS OF CONTRACT (GCC)

1. Prices

- 1.1 Prices should be quoted only in FOR IIT Bhilai Basis and should be inclusive of all taxes and duties.
- 1.2 The statutory taxes and duties applicable at the time of supply of material shall be applicable and will be reimbursed on submission of requisite documents/ proofs as the relevant Government Notification.
- 1.3 **IIT Bhilai is eligible for concessional rate of GST i.e. @ 5% under the Notification No. 45/2017, Central Tax (Rate), State Tax (Rate) and/or 47/2017, Integrated Tax (Rate), dated 14.11.2017 subject to applicability of the cited clause of GST on items based on directives of DSIR.**
- 1.4 In case due to any error/ oversight, the GST quoted by the bidder is less than the actual rate as per tariff, the bidder will not be permitted to rectify the error/oversight. The orders/ contract will be placed for the total amount including the (lower) rate/s quoted by the bidder, with reduced basic amount to the extent of difference in tax/duty amount, so that the total amount (basic + actual rate as per tariff), remains same(quoted basic + quoted rate). The difference amount payable, if any, between the quoted rate and actual rate as per tariff (without exemption) shall be borne by the bidder. The exemption certificate, if applicable, will be issued for the reduced basic amount only.

2. Software Licenses

The software licenses, if any, shall be required in the name of the user. The licenses shall contain paper licenses and at least one set of media (CDs) – wherever applicable. All software licenses quoted should be of perpetual.

3. Performance Bank Guarantee (PBG)

The successful bidder has to submit performance security for an amount of 3% of the order value in the form of BG from any nationalized bank within 14 (Fourteen) days from the date of award of contract and should be valid for 60 days beyond the date of completion of all contractual obligation including warranty obligation. And it will be returned after 60 days beyond the warranty period.

4. Completeness Responsibility

Notwithstanding the scope of work, engineering, supply and services stated in bid document, any equipment or material, engineering or technical services which might not be even specifically mentioned under the scope of supply of the bidder and which are not expressly excluded there from but which – in view of the bidder - are necessary for the performance of the equipment in accordance with the specifications are treated to be included in the bid and has to be performed by bidder. The items which are over & above the scope of supply specified in the Schedule of Requirements may be marked as “Optional Items”.

5. Warranty

All the equipment and components supplied must have 3 (Three) year warranty covering all parts & labour starting from the date after the successful installation and commissioning. During the warranty period, Bidder will have to undertake comprehensive maintenance of the entire hardware components, software, equipment, support and accessories supplied by the Bidder at the place of installation of the equipment. It is preferable that additional 2-years warranty is provided as an optional item. All site visits during the warranty period must be covered by the vendor/OEM at their own expense. Unlimited telephone and remote support for diagnostics and troubleshooting during instrument lifetime should be provided. The Supplier warrants that all the Goods are new, unused, and of the most recent or current models, and that they incorporate all recent improvements in design and materials, unless provided otherwise in the Contract. The supplier further warrants that all Goods supplied under this contract shall have no defect arising from design, materials or workmanship (except when the design and/or material is required by the Purchaser’s specifications) or from any act or omission of the supplier. The warranty should be comprehensive on site, repair/replacement basis free of cost. The defects, if any, during the guarantee/warranty period are to be rectified free of charge by arranging free replacement wherever necessary. Goods requiring

Supply and Installation of High Temperature Tribometer at IIT Bhilai

warranty replacements must be replaced on a free of cost basis. Collecting of faulty Hardware from onsite and provisioning the replacement of Hardware under warranty at onsite shall be the responsibility of the supplier. IIT Bhilai reserves the right to invoke the Performance Bank Guarantee submitted by bidder, in case a) Supplied equipment & Software components fail to achieve the performance as stipulated in this document and / or b) The bidder fails to provide the warranty and other services in scheduled time frame, as stipulated in this document, penalty clauses mentioned under point 9 of this tender shall be applicable.

6. Training:

Suppliers need to provide Free of Charge onsite training on applications and maintenance should be provided to at least five persons for a minimum of five days by the certified company engineers. IIT Bhilai will not bear any training or living expenditure in this regard. The demonstration and training shall include different numbers of bulk materials as well as coated (thin and thick) samples prepared by the user. The Supplier should arrange for regular visits to the IIT Bhilai campus by its technical team and assist in the maintenance of the item/equipment within warranty period.

7. Payments

7.1. For INR Payment by NEFT – 100% against delivery, Installation, commissioning, acceptance by IIT Bhilai and submission of Performance bank Guarantee within 30 Days from the date of acceptance by IIT Bhilai.

7.2. Payments towards AMC shall be paid at the end of each quarter/ year.

8. Ship to

Billing or Delivery Address:

Construction of the permanent campus for IIT Bhilai is in process and expected to be completed by the end of October 2022. The complete setup should be delivered at the permanent campus if the construction of the permanent campus is completed, otherwise the set up should be delivered and installed at the temporary campus of IIT Bhilai. Supplier should deliver the complete setup as per the instructions of IIT Bhilai.

Transit/ temporary Campus address (Invoice should be raised in the name of):

The Director,
Indian Institute of Technology Bhilai,
GEC Campus, Sejbahar
Raipur – 492 015 Chhattisgarh, INDIA

Permanent Campus Address:

The Director,
Indian Institute of Technology Bhilai,
Kutelabhata
Khapri, Chhattisgarh 491001, INDIA

9. Liquidated Damage

IIT Bhilai reserves the right to levy penalty @ of 0.5 % of order value per week of delay in supply and /or installation, beyond the schedule as mentioned in this tender document subject to maximum of 10% of the order value. IIT BHILAI reserves the right to cancel the order in case the delay is more than 06 weeks. The delay in delivery and/or installation not attributed to supplier viz. delay in site preparation, delay in submission of required documents etc. and the conditions arising out of Force Majeure will not be considered for the purpose of calculating penalties.

10. Force Majeure

IIT Bhilai may consider relaxing the penalty and delivery requirements, as specified in this document, if and to the extent that, the delay in performance or other failure to perform its obligations under

Supply and Installation of High Temperature Tribometer at IIT Bhilai

the contract is the result of a Force Majeure. Force Majeure is defined as an event of effect that cannot reasonably be anticipated such as acts of God (like earthquakes, floods, storms etc.), acts of states / state agencies, the direct and indirect consequences of wars (declared or undeclared), hostilities, national emergencies, civil commotion and strikes at successful Bidder's premises or any other act beyond control of the bidder.

11. Arbitration

All disputes/claims of any kind arising out of supply, commissioning, acceptance, warranty maintenance etc. under this Contract, shall be referred by either party (IIT Bhilai or the bidder) after issuance of 30 days' notice in writing to the other party clearly mentioning the nature of dispute to the Sole Arbitrator appointed by IIT Bhilai. The venue for arbitration shall be specified in the purchase order/agreement. The arbitration proceedings shall be conducted in English and as per the provisions of Indian Arbitration and Conciliation Act, 1996. The decision of the Arbitrator shall be final and binding on both the parties.

12. Jurisdiction

The disputes, legal matters, court matters, if any shall be subject to Raipur jurisdiction only.

13. Indemnity

Selected bidder shall save, indemnify and hold harmless IIT Bhilai from any third-party Govt. Claims, losses penalties, if any, arising in connection with this Contract.

14. Assignment

Selected bidder/ Party shall not assign, delegate or otherwise deal with any of its rights or obligations under this Contract without prior written permission of IIT Bhilai.

15. Severability

If any provision of this Contract is determined to be invalid or unenforceable, it will be deemed to be modified to the minimum extent necessary to be valid and enforceable. If it cannot be so modified, it will be deleted and the deletion will not affect the validity or enforceability of any other provision.

16. Unloading

The bidder should ensure to unload all the materials (equipment and other related accessories etc) to the installation site at IIT Bhilai (First floor of the academic building). Unloading of the goods at IIT Bhilai is the responsibility of the bidder. No manpower will be provided by IIT Bhilai.

17. Relocation to Permanent Campus:

The bidder has to stand guaranteed for the relocation of the equipment once the permanent campus of IIT Bhilai gets ready for shifting. This is the bidder's responsibility to dismantle the set up in temporary campus and re-install it to our Kutelabhata campus around 45 km away. The dismounting, packing, insurance, transport, material-handling system support, unpacking, re-installation and commissioning of the system with test running and its conformity must be carried out by the bidder. Cost for the same, if any, shall be quoted separately-

***** (END OF SECTION IV) *****

Supply and Installation of High Temperature Tribometer at IIT Bhilai
SECTION – V: SCHEDULE OF REQUIREMENTS

S.No.	Particulars of the item required	Qty Required
1	High Temperature (1000 °C) Tribometer	1 No

Technical Specification for High Temperature (1000 °C) Tribometer

S.No.	Technical Specifications:	
1.	Type of Machine	<p>The assembly is expected to be constituted by the following modules (i) mechanical components for providing motion and support, (ii) electrical components, (iii) a furnace, and (iv) cooling system.</p> <p>The instrument should be able to perform wear and friction testing of solid materials in lubricated or dry conditions with operating temperature up to 1000 °C as per ASTM G99 or similar international standard. The instrument should be supported by a software that is built to operate, make measurements, display real time data, store the data and aid in analysis of the results. The tribometer should be designed to use water as coolant to protect the mechanical parts during high temperatures.</p> <p>It should be able to operate in both rotational and reciprocating mode.</p>
2.	Design and Construction	<ul style="list-style-type: none"> ● A tabletop model, compact, robust, made of high-quality parts and materials. ● Fully conform to standards ASTM G99 or DIN 50324. ● Must include both rotary and linear/angular reciprocation module of tribo-testing, both possible to operate in high temperature conditions. ● Having safety features built-in to protect the furnace, electronics, mechanical system, load cell etc. ● The layout should be such that the electronics are well separated from water lines and heat produced from the furnace. ● The environment in the chamber should not affect the electronics and mechanical systems ● The environment in the test chamber must not be contaminated by the air and/or water circulation ● It should be easily serviceable
3.	Motor	<ul style="list-style-type: none"> ● Torque: not less than 400 mN-m ● Speed: 0.2 RPM or less to 1500 RPM or more ● Capable of rotating in both directions ● Capable of angular reciprocation with a frequency 0.01-5 Hz ● Angular resolution of 0.1 degrees ● Stroke length for linear reciprocation module: 1–25mm or better.

Supply and Installation of High Temperature Tribometer at IIT Bhilai

4.	Mechanical system	<ul style="list-style-type: none"> ● The mechanical system (shaft, bearings, and the furnace linings etc.) should use superior high temperature materials ● It should be robust ● Easy to operate ● Controllable using software. ● Withstands high temperatures without oxidation and wear.
5.	Loading	<ul style="list-style-type: none"> ● Max. Load (N): 50 or Higher ● Min. Load (N): 1 or Lower ● Friction Force Resolution (mN):0.05 or better ● Minimize thermal drift error on friction measurement ● Should avoid mutual influence between normal load and friction force signals ● Should have highly linear and precise elastic arm for friction load measurement
6.	Heating system	<ul style="list-style-type: none"> ● Precise and homogeneous sample temperature control up to 1000 °C ● Max. Oven Temperature: At least 1000 °C ● Max. Sample Temperature: At least 1000 °C ● Uniform heating of the sample from all sides must be achievable.
7.	Electronics	<ul style="list-style-type: none"> ● The electronic circuit should be built using high quality electronics and shall be coated with suitable coatings to protect it from condensed water and climatic conditions. ● The electronic circuitry must have adequate cooling.
8.	Cooling	<ul style="list-style-type: none"> ● Chillers shall be provided with water cooling to the furnace walls and the mechanical system wherever necessary. ● The water lines must be sealed and far away from the electronics. ● There must be an in-built safety system to protect the furnace wall and linings during heating if the water flow rates are low. ● The water conduits should be capable of easy cleaning and maintenance.
9.	Software	<ul style="list-style-type: none"> ● Tribometer operation and data display software ● Continuous real time data acquisition of friction and wear depth signal during measurement. ● Windows 10 compatible. ● Continuous real time data acquisition of friction force signal during a test. ● Adjustable data acquisition frequency at least up to 400Hz.
10.	Automatic calibration procedures	<ul style="list-style-type: none"> ● Force measurement calibration; ● Speed calibration; ● First lap synchronization: each test can be started at the same position of the circle.

Supply and Installation of High Temperature Tribometer at IIT Bhilai

11.	Calibration	<p>User can perform in a very short time the three following calibrations:</p> <ul style="list-style-type: none"> ● Tangential force calibration; ● Rotational speed calibration; ● Radius sensor calibration.
12.	Accessories	<ul style="list-style-type: none"> ● stainless steel ball holder to accommodate a 10 mm ball (at least 2 pcs) ● Inconel ball holder to accommodate a 6 mm ball (at least 4 pcs) ● stainless steel pin holder to accommodate a 6 mm dia. pin (at least 2 pcs) ● Inconel ball holder to accommodate a 3 mm ball (at least 4 pcs) ● stainless steel pin holder to accommodate a 3 mm dia. Pin (at least 2 pcs) ● Inconel disk sample 2 inch (at least 2 pcs) (suitable for wear test above 850 degree C) ● Alumina disk 2 inch (at least 2 pcs) ● Universal sample holder (1 pc) of stainless steel (for room temperature test) to accommodate a maximum dimension of 2 inches x 2 inches x 10 mm (thickness). Samples of different sizes and shapes within this limit must be accommodated without any sacrifice in the test quality. ● Inconel sample holders to hold the samples (for high temperature tests) with the following dimensions <ol style="list-style-type: none"> 1. 20 mm x 20 mm x 5 mm thickness (2 pc) 2. 20 mm x 20 mm x 10 mm thickness (1 pc) 3. 20 mm diameter and 5 mm thickness (2 pc) 4. 20 mm diameter and 10 mm thickness (1 pc) ● Heat resistant glove (at least 2 sets) ● Various tool kits for operating on the tribometer
13.	Consumables	<ul style="list-style-type: none"> ● 10 mm diameter 100Cr6 balls, 10 pieces ● 6 mm diameter 100Cr6 balls, 100 pieces ● 6 mm diameter 100Cr6 pin, 10 pieces ● 3 mm diameter 100Cr6 balls, 100 pieces ● 6 mm diameter 100Cr6 pin, 10 pieces ● 6 mm diameter Al₂O₃ (alumina) balls, 50 pieces ● 6 mm diameter Al₂O₃ (alumina) pins, 10 pieces ● 3 mm diameter Al₂O₃ (alumina) balls, 20 pieces ● 3 mm diameter Al₂O₃ (alumina) pins, 10 pieces ● 6 mm diameter WC (hard metal) balls, 20 pieces ● 3 mm diameter WC (hard metal) balls, 20 pieces ● 6 mm diameter Inconel 718 balls, 20 pieces ● 6 mm diameter Inconel 625 balls, 20 pieces
14.	Testing features,	<ul style="list-style-type: none"> ● Pin on Disc and Ball on Disc

Supply and Installation of High Temperature Tribometer at IIT Bhilai

	Configurations and Add-ons	<ul style="list-style-type: none"> ● Sample and ball/pin holder for easy tribology testing setup for extremely high temperature
15.	Sensors and standards	<ul style="list-style-type: none"> ● Adjustable data acquisition rate up to 400 Hz for recording frictional phenomena with high fidelity. ● Fully conform to ASTM G99, ASTM G133 and DIN 50324 standards . ● Differential arm with two LVDT/piezoelectric sensors for precise temperature drift compensation . ● Radius sensor allows an automatic display of the arm on the sample. ● Motorized and software controlled tribometer arm used for normal loading. ● Carriage with graduated guide and adjustment knob . ● Online wear depth sensor. ● 2 Independent thermocouples for sample and oven; 1 additional thermocouple for TOP Heating Cover in case of temperature up to 1000 °C. ● Electrical contact resistance . ● Cell for lubricants. ● Plastic Cover: For low temperature measurement.
16.	In-built humidity and temperature sensor integrated with the software;	<ul style="list-style-type: none"> ● Temperature sensor range: -40 to 125 °C ● Temperature sensor resolution 0.015 °C ● Humidity sensor range 0 to 100 %RH ● Humidity sensor resolution 0.01 %RH ● Liquid Cup for tribology in liquid medium
17.	Dedicated Desktop Computer	Should be supplied along with the tribometer which will house the software for the tribometer. (Windows operating system, 32 GB ram, 1TB SSD storage, 27 inch monitor, i7 processor)
18.	Uninterrupted power supply (UPS)	10 kVA or better
Other mandatory requirements:		
19.	Certifications	<ul style="list-style-type: none"> ● Please provide documentation on High Temperature (1000 °C) Tribometer origin and OEM certification. ● Authorization certificate from all the original equipment manufacturers (OEM) must also include that in case of non-availability of the vendor, they will stand guaranteed for after-sale services.
20.	Pre-dispatch inspection	Pre dispatch inspection cost should be arranged at manufacturer's site for two days (to and fro return air fare, fooding and lodging) shall be borne by the Supplier if the OEM is from India. If the OEM is from a foreign country, the pre-dispatch inspection shall be carried

Supply and Installation of High Temperature Tribometer at IIT Bhilai

		out via video conferencing. All the points related to machine specifications and repeatability (upto $\pm 5\%$ deviation) must be demonstrated by the supplier during pre-dispatch inspection. All the measuring devices for the same purpose must be arranged by the supplier during demonstration. The pre-dispatch inspection shall include bulk materials as well as coated (thin and thick) samples prepared by the user.
21.	Calibration and maintenance	The time-bound or error-based calibrations of the data acquisition system is required and other maintenance of instruments need to be covered during complete warranty duration and done by the supplier free of cost.
22.	Transportation and other charges	The tender should be inclusive of all the charges till IIT Bhilai stores including the transportation, installation, maintenance, custom, freight, insurance, demurrage etc. and should be free from any hidden charge(s). During delivery of the system at IIT Bhilai, loading-unloading from the truck and placement of the consignment at the designated place must be undertaken by the supplier at their own cost.
23.	Insurance Coverage	The successful bidder should provide the insurance (free of cost) to the machine/equipment till IIT Bhilai at the time of delivery due to the sensitive nature of the Machine. They should take this into consideration while quoting their bid/quotation.

Note: Bidder must submit a technical compliance sheet (with bidder specification, deviation & justification) and all the relevant documents to ensure that the technical specifications are compiled & all the conditions are fulfilled. Note: For Technical Queries contact: purnendu@iitbhilai.ac.in

Authorized Signatory:
(Signature of the Bidder, with Official Seal)

***** (END OF SECTION V) *****

Supply and Installation of High Temperature Tribometer at IIT Bhilai

SECTION – VI PRICE BID

PRICE BID -Schedule of price bid in the form of BOQ format:

1. Financial Proposal/Commercial bid format is provided as BoQ along with this tender document at <https://eprocure.gov.in/eprocure/app>
2. Bidders are advised to download this BoQ.xls as it is and quote their offer/rates in the permitted column and upload the same in the commercial bid.
3. Bidder shall not tamper/modify downloaded price bid template in any manner. In case if the same is found to be tampered/modified in any manner, tender will be completely rejected and EMD shall be forfeited.
4. **Any element of cost, taxes, duties levies etc. not specifically indicated in the BOQ, shall not be paid by the purchaser. If GST amount is not quoted in the BOQ (Financial Bid), the total cost will be treated as inclusive of GST. No further communication will be entertained later or else the EMD will be forfeited.**
5. The tender shall remain valid for acceptance for 180 days, from the date of tender opening.

OTHER CONDITIONS FOR PRICE BIDS

1. No unilateral revision in price will be admissible.
2. Rates should be quoted in the accounting units (A/U) mentioned in this tender. Rates must be quoted clearly on a free delivery basis at IIT Bhilai and total value is also indicated in words.
3. Any optional indicated in techno-commercial bids must be priced separately.
4. In case spares/accessories are applicable, their list and price should be clearly indicated separately.
5. The price must be stated for each item separately. The percentage of reduction, in the unit price should also be quoted, should an order to that extent be placed with you.
6. The bids shall be evaluated on the basis of price quoted (Including GST) for Sl. No. 1 i.e. Basic Equipment price and Essential Accessories in the Commercial bid. Price quoted for the Sl. No. 2, 3 and 4 will be optional and will not be taken into consideration in deciding the L1 Bidder.
7. Any information related to essential accessories or optional accessories should be mentioned in the remarks column.

Signature of the Tenderer

Seal:

(END OF SECTION VI)

Supply and Installation of High Temperature Tribometer at IIT Bhilai

SECTION –VII ANNEXURES (STANDARD FORMATS)

ANNEXURE A

TENDER ACCEPTANCE LETTER

(To be submitted on Company Letter Head)

Date: _____

To,

**The Director
Indian Institute of Technology Bhilai,
GEC Campus Sejbahar,
Raipur 492015 Chhattisgarh**

SUB: Acceptance of Terms & Conditions of Tender.

Tender Reference No./ Date:

Name of Tender / Work:

Dear Sir,

1. I / We have downloaded / obtained the tender document(s) for the above mentioned 'Tender/Work' from the web site(s) namely: www.iitbhilai.ac.in, <http://eprocure.gov.in/eprocure/app> as per your NIT / advertisement, given in the above-mentioned website(s).
2. I / We hereby certify that I / We have read the entire terms and conditions of the tender documents (including all documents like annexure(s), schedules(s), etc.), which form part of the contract agreement and I / We shall abide hereby by the terms / conditions/ clauses contained therein.
3. The corrigendum(s) issued from time to time by your department / organization too has also been taken into consideration, while submitting this acceptance letter.
4. I / We hereby unconditionally accept the tender conditions of above mentioned tender document(s) / corrigendum(s) in its totality / entirety.
5. I / We do hereby declare that our Firm has not been blacklisted / debarred by any Govt. Department/Public sector undertaking.
6. I / We do hereby declare that our firm will submit the Performance Bank Guarantee for 3% PO Value within two weeks from the date of award of contract, if Contract is awarded to our firm.
7. I / We certify that all information furnished by our Firm is true & correct and in the event that the information is found to be incorrect/untrue or found violated, then your department/ organization shall without giving any notice or reason therefore or summarily reject the bid or terminate the contract, without prejudice to any other rights or remedy including the forfeiture of the fully said earnest money deposit absolutely.

Yours Faithfully,

Authorized Signatory:
(Signature of the Bidder, with Official Seal)

ANNEXURE B

Supply and Installation of High Temperature Tribometer at IIT Bhilai

AUTHORITY LETTER

AUTHORIZATION

We _____ (name of the bidder) hereby authorise Shri / Smt. _____ (name of the authorised person) to sign and submit the bid to IIT Bhilai against their Tender No. _____ dated _____. Shri / Smt. _____ (name) is also authorised to negotiate the terms and conditions pertaining to the said tender on behalf of M/s _____ (name of bidder). The specimen signature of Shri / Smt. _____ (name) is appended below.

Specimen Signature:

Name:

The undersigned is authorised to delegate the authority on behalf of M/s _____ (name of bidder), as stipulated above

For _____
(name of bidder)

Signature & Company Seal:

Name:

Designation:

Email:

Mobile No.:

Supply and Installation of High Temperature Tribometer at IIT Bhilai
ANNEXURE C

SUMMARY INFORMATION SHEET

Sr. No.	Information heading	Details
1	Name of the Vendor	
2	Name of the person(s) to whom, all references shall be made, regarding this TENDER/PROJECT (Authorized Signatory)	
3	Address of the Contact Person	
4	Designation of the Contact Person	
5	Telephone number(s) of the Contact Person.	
6	Mobile number of the Contact Person	
7	Fax number of the Contact Person	
8	Email ID(s) of the Contact Person	
9	Annual Turnover (in Lakhs): Last three Years (In INR)	1) 2) 3)
10	Years of Operation (in Years):	
11	PAN Number	
12	GSTN	
13	Bank Name and Account No.	
14	IFS Code	

Authorized Signatory:
(Signature of the Bidder, with Official Seal)

Supply and Installation of High Temperature Tribometer at IIT Bhilai
ANNEXURE D

PROFORMA OF BANK GUARANTEE
(on non-judicial paper of appropriate value)

To,

The Director
Indian Institute of Technology Bhilai,
GEC Campus, Old Dhamtari Road
Sejbahar, Raipur 492015 Chhattisgarh

BANKS GUARANTEE NO:

DATE:

Dear Sir(S)

This has reference to the Purchase Order No. ___ Dated _____ been placed by Indian Institute of Technology Bhilai (IIT Bhilai), Raipur on M/s _____ (Name & Address of Bidder) for supply, installation, commissioning and warranty of _____ (description of items) at client's site.

The conditions of this order provide that the Bidder shall,

1. Arrange to deliver the items listed in the said order to the consignee, as per details given in said order, and
2. Arrange to install and commission the items listed in said order at client's site, to the entire satisfaction of IIT Bhilai and
3. Arrange for the comprehensive warranty service support towards the items supplied by Bidder on site as per the warranty clause in said purchase order.

M/s (Name of bidder) has accepted the said purchase order with the terms and conditions stipulated therein and have agreed to issue the performance bank guarantee on their part, towards promises and assurance of their contractual obligations vide the Supply Order No. ___ M/s. ___ (name of vendor) holds an account with us and has approached us and at their request and in consideration of the promises, we hereby furnish such guarantees as mentioned hereinafter.

IIT Bhilai shall be at liberty without reference to the Bank and without affecting the full liability of the Bank hereunder to take any other undertaking of security in respect of the suppliers obligations and / or liabilities under or in connection with the said contract or to vary the terms vis-a – vis the supplier or the said contract or to grant time and or indulgence to the supplier or to reduce or to increase or otherwise vary the prices or the total contract value or to forebear from enforcement of all or any of the obligations of the supplier under the said contract and/or the remedies of IIT Bhilai under any security now, or hereafter held by IIT Bhilai and no such dealing(s) with the supplier or release or forbearance whatsoever shall have the effect of releasing the bank from its full liability of IIT Bhilai hereunder or of prejudicing right of IIT Bhilai against the bank.

This undertaking guarantee shall be a continuing undertaking guarantee and shall remain valid and irrevocable for all claims of IIT Bhilai and liabilities of the supplier arising up to and until ___(date).

This undertaking guarantee shall be in addition to any other undertaking or guarantee or security whatsoever the that IIT Bhilai may now or at any time have in relation to its claims or the supplier's obligations/ liabilities under and / or in connection with the said contract and IIT Bhilai shall have the full authority to take recourse to or enforce this undertaking guarantee in preference to the other undertaking or security(ies) at its sole discretion and no failure on the part of IIT Bhilai in enforcing or requiring enforcement of any other undertaking or security shall have the effect of releasing the bank from its full liability hereunder.

Supply and Installation of High Temperature Tribometer at IIT Bhilai

We ____ (Name of Bank) hereby agree and irrevocably undertake and promise that if in your (IIT Bhilai's) opinion any default is made by M/s ____ (Name of bidder) in performing any of the terms and /or conditions of the agreement or if in your opinion they commit any breach of the contract or there is any demand by you against M/s ____ (Name of bidder), then on notice to us by you, we shall on demand and without demur and without reference to M/s ____ (Name of bidder), pay you, in any manner in which you may direct, the amount of Rs. _____/- (Rupees _____ Only) or such portion thereof as may be demanded by you not exceeding the said sum and as you may from time to time require. Our liability to pay is not dependent or conditional on your proceeding against M/s ____ (Name of bidder) and we shall be liable & obligated to pay the aforesaid amount as and when demanded by you merely on an intimation being given by you and even before any legal proceedings, if any, are taken against M/s ____ (Name of bidder)

The Bank hereby waives all rights at any time inconsistent with the terms of this undertaking guarantee and the obligations of the bank in terms hereof shall not be anyway affected or suspended by reason of any dispute or disputes having been raised by the supplier (whether or not pending before any arbitrator, Tribunal or Court) or any denial of liability by the supplier or any order or any order or communication whatsoever by the supplier stopping or preventing or purporting to stop or prevent payment by the Bank to IIT Bhilai hereunder.

The amount stated in any notice of demand addressed by IIT Bhilai to the Bank as claimed by IIT Bhilai from the supplier or as suffered or incurred by IIT Bhilai on the account of any losses or damages or costs, charges and/or expenses shall as between the Bank and IIT Bhilai be conclusive of the amount so claimed or liable to be paid to IIT Bhilai or suffered or incurred by IIT Bhilai, as the case may be and payable by the Bank to IIT Bhilai in terms hereof.

You (IIT Bhilai's) shall full liberty without reference to us and without affecting this guarantee, postpone for any time or from time to time the exercise of any of the powers and rights conferred on you under the contract with the said M/s ____ (Name of bidder) and to enforce or to forbear from endorsing any power or rights or by reason of time being given to the said M/s ____ (name of bidder) which under law relating to the sureties would but for the provisions have the effect of releasing us.

You will have full liberty without reference to us and without affecting this guarantee, postpone for any time or from time to time the exercise of any of the powers and rights conferred on you under the contract with the said M/s ____ (Name of bidder) and to enforce or to forbear from endorsing any power or rights or by reason of time being given to the said M/s ____ (Name of bidder) which under law relating to the sureties would but for the provisions have the effect of releasing us.

Your right to recover the said sum of Rs. ____/- (Rupees ____ only) from us in manner aforesaid will not be affected/ or suspended by reason of the fact that any dispute or disputes have been raised the said M/s ____ (Name of bidder) and/ or that any dispute or disputes are pending before any officer, tribunal or court or Arbitrator.

The guarantee herein contained shall not be determined or affected by the liquidation or winding up, dissolution or change of constitution or insolvency of the said M/s ____ (Name of bidder) but shall in all respects and for all purposes be binding and operative until payment of all dues to IIT Bhilai in respect of such liability or liabilities.

Our liability under this guarantee is restricted to Rs. _____/- (Rupees _____ Only). Our guarantee shall remain in force until unless a suit action to enforce a claim under guarantee is filed against us within six months from (which is date of expiry of guarantee) all your rights under the said guarantee shall be forfeited and we shall be relieved and discharged from all liabilities there under.

We have power to issue this guarantee in your favour under Memorandum and Articles of Association of our Bank and the undersigned has full power to do under the power of Attorney dated.

Supply and Installation of High Temperature Tribometer at IIT Bhilai

Notwithstanding anything contained herein:

- a) Our liability under this guarantee shall not exceed Rs____(in words)
- b) This bank guarantee shall be valid up to_& unless a suit for action to enforce a claim under guarantee is filed against us within six months from the date of expiry of guarantee. All your rights under the said guarantee shall be forfeited and we shall be relieved and discharged from all liabilities there after i.e. after six months from the date of expiry of this Bank guarantee
- c) We are liable to pay the guaranteed amount or any parts thereof under this bank guarantee only and only if you serve upon us a written claim or demand or before _____
- d) The Bank guarantee will expire on Min 36 months for the warranty period + 2 Months for Claim period from the date of successful installations of the items in the order) _____Granted by the Bank

Signature of the authorized officer of the bank

Name of the officer_____

Designation of the officer_____

Seal, name, email and address of the Bank and address of the Branch.

Supply and Installation of High Temperature Tribometer at IIT Bhilai

ANNEXURE E

PREVIOUS WORK ORDER DETAILS

(To be submitted on Company Letter Head)

Evaluation Criteria	Name of the Client	Order No. & Date	Amount	Remark
List of Purchase Order / Work Orders where the similar systems (as per the technical specifications mentioned in the present tender) as per the Section-III	1.			Supporting document are to be attached
	2.			
	3.			
	4.			
	5.			

Date:

Place:


Authorized Signatory

Seal

Name:

Designation:

Contact No:



Supply and Installation of High Temperature Tribometer at IIT Bhilai

ANNEXURE F

TURNOVER DETAILS

(To be submitted on Company Letter Head)

Evaluation Criteria			Remark
Bidder's Annual Turnover for last three financial years	Financial Year	Turnover in Rs.	
	2020-21		Supporting Documents are to be attached.
	2019-20		
	2018-19		

Authorized Signatory:
(Signature of the Bidder, with Official Seal)

Supply and Installation of High Temperature Tribometer at IIT Bhilai

ANNEXURE G

Form - A

NO PURCHASE PREFERENCE - SELF DECLARATION

(TO BE GIVEN BY VENDORS WITH BID)

NIT No.	To be filled by vendor
Tendered Item/ Equipment	To be filled by vendor

With reference to the above mention NIT, we hereby confirm that for this case **we do not wish to avail the Preference to Make in India** as the made in India content / indigenous content as a **percentage of the Basic Cost** of the Product / Service is less than 50%.

Indigenous %	Please Mention
Imported %	Please Mention
Name of firm	
Address	
Details of technology collaboration agreement/ transfer of technology agreement, if any.	

We undertake that any False declaration on our part is a breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which we or our successors will be debarred for up to 2 years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under the law.

Signature with seal

Supply and Installation of High Temperature Tribometer at IIT Bhilai

Form – B

SELF DECLARATION OF LOCAL CONTENT AND PURCHASE PREFERENCE

(TO BE GIVEN BY VENDORS WITH BID)

NIT No.	To be filled by vendor
Tendered Item/ Equipment	To be filled by vendor

With reference to the above mention NIT, we hereby confirm that for this case we wish to avail the Preference to Make in India as made in India content / indigenous content as a percentage of the Basic Cost of the Product / Service as given below:		
Specify Indigenous %	Please Mention	
Imported %	Please Mention	
Classification	Class – I Supplier	Class – II Supplier
Name of Indian Partnering firm with document or consent letter		
Details of Location(s) of local value addition claimed as indigenous		

Note: Bidder will be required to submit a cost auditor certificate (on percentage of local content) for the procurement value more than 10 Crores along with this declaration.

We undertake that any False declaration on our part is a breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which we or our successors will be debarred for up to 2 years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under the law.

ANNEXURE H

Bid-Security Declaration Form

(To be submitted on Company Letter Head)

Date: _____

To:

**The Director,
Indian Institute of Technology Bhilai,
GEC Campus Sejbahar,
Raipur 492015 Chhattisgarh**

Tender Reference No./ Date:

Name of Tender / Work:

Dear Sir,

We, the undersigned, declare that:

We understand that, according to your conditions, bids must be supported by a Bid-Security Declaration.

We accept that we will automatically be suspended from being eligible for bidding in any contract with the **Indian Institute of Technology (IIT) Bhilai** for the period of time of **3 years** starting from February 2022 , if we are in breach of our obligation(s) under the bid conditions, because we:

- (a) have withdrawn our Bid during the period of bid validity specified in the Letter of Bid; or
- (b) having been notified of the acceptance of our Bid by the **Indian Institute of Technology (IIT) Bhilai** during the period of bid validity, (i) fail or refuse to execute the Contract, if required, or (ii) fail or refuse to furnish the Performance Security, in accordance with the terms and conditions mentioned in the tender document .

We understand this Bid-Security Declaration shall expire if we are not the successful Bidder, upon the earlier of (i) our receipt of your notification to us of the name of the successful Bidder; or (ii) twenty-eight days after the expiration of our Bid.

Authorized Signatory:

(Signature of the Bidder, with Official Seal)

Supply and Installation of High Temperature Tribometer at IIT Bhilai

CHECK LIST - UNDERTAKING

S.NO	Document Description	Enclosed (Yes/No)	Page No.	Remarks
1	Scanned copy of Tender fee payment advice of Rs.1500/- showing DU number clearly or valid MSME/NSIC Exemption certificate.			
2	EMD: fee payment advice of Rs.2,60,000/- showing DU number clearly Valid MSME/NSIC Exemption certificate along with Bid Security declaration			
3	Scanned copy of OEM/ Authorized dealer certificate			
4	Details of Similar setups supplied as per Pre-eligibility			
5	Detailed Price bid			
6	Annual turnover details for last three years			
7	Scanned signed copies of Annexures from A to G, whichever is applicable			
8	Scanned copy of Technical Compliance			

I / We do hereby declare that all the above mentioned documents are enclosed as per the tender document.

Authorized Signatory:
(Signature of the Bidder, with Official Seal)