



भारतीय प्रौद्योगिकी संस्थान भिलाई
जी.ई.सी. कैंपस, सेजबहार, रायपुर - ४९२०१५
छत्तीसगढ़, भारत

Indian Institute of Technology Bhilai
G.E.C. Campus, Sejbahar, Raipur - 492015
Chhattisgarh, India

IITBh/Goods/CIF/2021-22/051

Date: 19.07.2021

CORRIGENDUM-1

It is notified to all concerned parties that with reference to our Tender No. IITBh/Goods/CIF/2021-22/051 dated 05.07.2021 for the “Supply and Installation of High Resolution Transmission Electron Microscope with accessories at IIT Bhilai”, the following changes are made in the tender document:

The following specifications are changed in the Technical Specifications of “Supply and Installation of High Resolution Transmission Electron Microscope with accessories at IIT Bhilai”:

S.No.	Page No./Section Features/ (Clause)	Detailed specification of the features		
		Item	Existing	Modification/Correction
1.	Section-V Page-17, Sl. No. 14 Point No. (c)	LN2 Free EDS	c) The EDS resolution must be ≤ 129 eV Mn K α .	c) The EDS resolution must be ≤ 136 eV Mn K α .
2.	Section-V Page-17, Sl. No. 14 Point No. (g)	LN2 Free EDS	g) The EDS software must be operated through the same Graphics user interface of the camera, for ease of use.	Deleted
3.	Section-V Page-19 Sl. No. 21 Sub-Point	Site preparation and environmental conditions	Supplier or their authorized agent needs to verify/survey the installation site prior to submitting the tender bids. It is the supplier's responsibility to prepare the site as required for the system to deliver the quoted specifications. In addition, it is the supplier's responsibility to maintain the temperature and humidity of the site. This includes complete precision air-conditioning with all supporting equipment, such as magnetic canceller and anti-vibration devices if required. Nitrogen cylinder with regulator, liquid nitrogen and other	Supplier or their authorized agent needs to verify/survey the site prior to installation. The supplier shall support the site preparation activity as per the requirement of the instrument. This includes complete precision air-conditioning (which shall be provided by the institute) with all supporting equipment (to be provided by the vendor), such as magnetic canceller and anti-vibration devices if required. Nitrogen cylinder with regulator, liquid nitrogen and other consumables, including but not limited to electrical switches,

			consumables, including but not limited to electrical switches and whatever is required during installation shall be provided by the vendor without any extra cost.	standard samples to demonstrate complete functionalities, etc., and whatever is required for installation, testing and commissioning shall be provided by the vendor without any extra cost.
--	--	--	--	--

All other terms and conditions of the tender remain unchanged.


Special Administrative Officer