




CORRIGENDUM-1

It is notified to all concerned parties that with reference to our Tender No. IITBh/Goods/Chem/2021-22/105 dated 09.09.2021 for the "Supply, Installation and Commissioning of Automated Column Chromatography for Molecular Separation with Accessories at IIT Bhilai", the following changes are made in the tender document:

Addition or Changes in the Technical Specifications of "Supply, Installation and Commissioning of Automated Column Chromatography for Molecular Separation with Accessories at IIT Bhilai" The remaining technical specifications were not accepted for further change.			
S.No.	Page No./Section Features/ (Clause)	Detailed specification of the features	
		Existing specification Description	Modification/Correction
1.	Section-V, Page-13-14, Serial No.- 07	Accessories: <ul style="list-style-type: none"> • All necessary accessories for the operation of the system should be included free of cost. • The system should be provided with all necessary tubing, etc. for installation and smooth running of the instrument. • Prepacked silica cartridges of following sizes and numbers should be supplied along with the system: <ul style="list-style-type: none"> ○ a) 4g to 5g: 20 Nos or more. ○ b) 10g to 12g: 20 Nos or more. ○ c) 20 g to 30 g: 15 Nos or more ○ d) 40 g to 50 g: 15 Nos or more ○ e) 80 g to 100 g: 10 Nos or more • Empty loader cartridges of following sizes and numbers should be supplied along with the system: <ul style="list-style-type: none"> ○ a) 10 g to 12 g: 40 Nos or more. ○ b) 20 g to 30 g: 30 Nos or more ○ c) 40 g to 50 g: 30 Nos or more ○ d) 80 g to 100 g: 20 Nos or more ○ e) 120g to 150 g: 10 Nos or more 	<ul style="list-style-type: none"> • All necessary accessories for the operation of the system should be included free of cost. • The system should be provided with all necessary tubing, etc. for installation and smooth running of the instrument. • Prepacked silica cartridges of following sizes and numbers should be supplied along with the system: <ul style="list-style-type: none"> ○ a) 4g to 5g: 20 Nos or more. ○ b) 10g to 12g: 18 Nos or more. ○ c) 20 g to 30 g: 12 Nos or more ○ d) 40 g to 50 g: 12 Nos or more • Empty loader cartridges of following sizes and numbers should be supplied along with the system: <ul style="list-style-type: none"> ○ a) 10 g to 12 g: 20 Nos or more. ○ b) 20 g to 30 g: 40 Nos or more ○ c) 40 g to 50 g: 20 Nos or more ○ d) 80 g to 100 g: 10 Nos or more ○ e) 120g to 150 g: 10 Nos or more

	<ul style="list-style-type: none"> ● Vendor should supply either (A) or (B) (A) C-18 columns with all available, variable sizes, in each size. ■ a) 20 g to 30 g: 2 Nos or more ■ b) 40 g to 50 g: 2 Nos or more (B) Empty Stainless Steel / Glass columns of following sizes and numbers should be supplied along with the system: ■ a) 10 g to 12 g: 2 Nos or more ■ b) 40 g to 50 g: 2 Nos or more ● All required filters should be provided with the cartridges. ● Silica gel (200-400 mesh size) 25 Kg - two units. 	<ul style="list-style-type: none"> ● Vendor should supply either (A) or (B) (A) C-18 columns with all available, variable sizes, in each size. ■ a) 20 g to 30 g: 2 Nos or more ■ b) 40 g to 50 g: 2 Nos or more (B) Empty Stainless Steel / Glass columns of following sizes and numbers should be supplied along with the system: ■ a) 10 g to 12 g: 2 Nos or more ■ b) 40 g to 50 g: 2 Nos or more ● All required filters should be provided with the cartridges. ● Silica gel (200-400 mesh size) 25 Kg - two units.
--	--	--

All other terms and conditions of the tender remain unchanged.


 23/09/2021
 Special Administrative Officer